



## SVP Test and Calibration certificate

SVP Type :	SVP70
SVP Serial No.	0811104

Date of issue : 18-02-2013

Temperature Calibration :	Hart 1504 s/n A6B554 & Thermistor s/n 3014
Point 1:	4.7 °C
Point 2:	16.6 °C
Point 3:	25.6 °C
Pressure Calibration :	Custom Built Tank (TestUnit ASF150 Ser# 41-10-0007-R03)
Point 1:	0 Bar
Point 2:	302.9 Bar
Point 3:	595.8 Bar

### RMS Speed of Sound Errors


Temperature Validation :	0.0119 m/s
Pressure Validation :	0.0267 m/s

Calibration & Final Function Test : Sign : Jes H. H. Pedersen

QA Signature : Inits : Oskunfa  
2013/02.21



RESON A/S, Fabriksvænget 13, DK-3550 Slangerup  
Fax: +45 4738 0066, Phone: +45 4738 0022



SVP 70 Ti-6000m  
EM7211 S/N 0811104



PN 900-63-1000-20



