



SEA-BIRD ELECTRONICS, INC.

13431 NE 20th St. Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

Pressure Test Certificate

Customer Atlantic Marine Center

Job Number 74589

Date 6/18/2013

Technician PS

Serial Number 054928

Low Pressure (PSI) 50 PSI

Time (Minutes) 15 Minutes

High Pressure (PSI) 10000 PSI

Time (Minutes) 45 Minutes

Pass ☒

Fail ☐

Comments

O rings were replaced.

