



Sea-Bird Scientific
 13431 NE 20th Street
 Bellevue, WA 98005
 USA

+1 425-643-9866
 seabird@seabird.com
 www.seabird.com

SENSOR SERIAL NUMBER: 3455
 CALIBRATION DATE: 12-Jan-18

SBE 4 CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.00854904e+001 CPcor = -9.5700e-008 (nominal)
 h = 1.55882337e+000 CTcor = 3.2500e-006 (nominal)
 i = -1.53964843e-003
 j = 2.08981321e-004

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (kHz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
0.0000	0.0000	0.00000	2.54570	0.00000	0.00000
-0.9999	34.6055	2.78918	4.93981	2.78915	-0.00003
1.0000	34.6057	2.95967	5.04943	2.95970	0.00003
15.0001	34.6056	4.24850	5.81093	4.24853	0.00003
18.5000	34.6045	4.59329	5.99815	4.59329	0.00000
29.0001	34.5943	5.67005	6.54810	5.66998	-0.00007
32.5001	34.5796	6.03938	6.72634	6.03942	0.00005

f = Instrument Output (kHz)

t = temperature (°C); p = pressure (decibars); δ = CTcor; ε = CPcor;

Conductivity (S/m) = (g + h * f² + i * f³ + j * f⁴) / 10 (1 + δ * t + ε * p)

Residual (Siemens/meter) = instrument conductivity - bath conductivity

