



Sea-Bird Scientific
 13431 NE 20th Street
 Bellevue, WA 98005
 USA

+1 425-643-9866
 seabird@seabird.com
 www.seabird.com

SENSOR SERIAL NUMBER: 3470
 CALIBRATION DATE: 16-Jan-19

SBE 4 CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.96488793e+000
 h = 1.56790351e+000
 i = -2.48837526e-003
 j = 2.87899869e-004

CPcor = -9.5700e-008 (nominal)
 CTcor = 3.2500e-006 (nominal)

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (kHz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
0.0000	0.0000	0.00000	2.52460	0.00000	0.00000
-1.0000	34.6408	2.79175	4.92370	2.79178	0.00003
1.0000	34.6414	2.96244	5.03331	2.96241	-0.00003
15.0000	34.6382	4.25207	5.79466	4.25207	-0.00000
18.5000	34.6361	4.59704	5.98177	4.59704	0.00000
29.0000	34.6268	5.67476	6.53149	5.67477	0.00001
32.5000	34.6117	6.04433	6.70949	6.04433	-0.00001

f = Instrument Output (kHz)

t = temperature (°C); p = pressure (decibars); δ = CTcor; ϵ = CPcor;

Conductivity (S/m) = $(g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

