



**SVP Test and Calibration certificate**

SVP Type :	SVP70
SVP Serial No.	2410039

Date of issue :	29-09-2010
-----------------	------------

Functionality Test : Sign : Jind Petersen

Temperature Calibration :	Hart 1504 s/n A6B554 & Thermistor s/n 3014
Point 1:	4.6 °C
Point 2:	16.6 °C
Point 3:	25.5 °C
Pressure Calibration :	Custom Built Tank (TestUnit ASF150 Ser# 41-10-0007-R03)
Point 1:	0 Bar
Point 2:	300.7 Bar
Point 3:	601 Bar

	<u>RMS Speed of Sound Errors</u>
Temperature Validation :	0.0204 m/s
Pressure Validation :	0.1424 m/s

Calibration Completed : Sign : Jind Petersen

Final Function Test : Sign : Jind Petersen

QA Signature : Inits : 



RESON A/S, Fabriksvangen 13, DK-3550 Slangerup  
Fax: +45 4738 0066, Phone: +45 4738 0022



SVP 70 TI-6000m  
EM7211 SIN:241039

PSM

SVP 70 Ti-6000m

EM7211 S/N:2410039